



Ref. Certif. No.

DE 2-023834

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Electrostimulation - device

Name and address of the applicant

miha bodytec GmbH  
Siemensstr. 1  
86368 Gersthofen, Deutschland

Name and address of the manufacturer

miha bodytec GmbH  
Siemensstr. 1  
86368 Gersthofen, Deutschland

Name and address of the factory

miha bodytec GmbH  
Siemensstr. 1  
86368 Gersthofen, Deutschland

Ratings and principal characteristics

power supply: AC 100-240 V, 50-60 Hz, 1.5 A  
control unit: DC 15-19 V, 1.4A (19V) -1.7 A (15 V)

Trademark (if any)

miha bodytec

Customer's Testing Facility (CTF) Stage used

N/A

Model / Type Ref.

miha bodytec II medical

Additional information (if necessary may also be reported on page 2)

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005+A1  
IEC 60601-1-6:2010+A1  
IEC 60601-1-11:2015  
IEC 60601-2-10:2012+A1  
for national deviations see test report

As shown in the Test Report Ref. No. which forms part of this Certificate

21277914 001

This CB Test Certificate is issued by the National Certification Body



TÜVRheinland®

TÜV Rheinland LGA Products GmbH  
Tillystraße 2 · 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3936  
Mail: cert-validity@de.tuv.com  
Web: www.tuv.com



Date: 02.10.2018

Signature: